

Aldevron Certificate of Analysis Lot MPM027-03

Date of Manufacture: 06-AUG-2018



5602 Research Park Blvd
 Madison, Wisconsin 53719, USA
 phone: 1 (701) 297-9256
 toll-free: 1 (877) 787-3362
 fax: 1 (608) 441-3460
 email: protein@aldevron.com

Protein Name:	Item Code:
sNLS-SpCas9-sNLS Nuclease	9212-5MG
Storage Temperature:	Final Buffer:
-10 to -30 °C	25 mM Tris-HCl, 300 mM NaCl, 0.1 mM EDTA, 50% (v/v) Glycerol, pH 7.4
Container Fill:	Usage:
5.0 mg (0.50 mL)	Research use only

Assay	Method	Specifications	Results
Purity, RP-HPLC	RP-HPLC method	≥95%	Pass
Purity, SDS-PAGE	Densitometry SYPRO® Orange	≥95%	Pass
Aggregation	Analytical SEC	≤5% aggregate	Pass
Concentration	UV VIS 280 nm	10.0 mg/mL +/- 10%	Pass
Endotoxin	Kinetic turbidimetric LAL	≤ 10 EU/mg	Pass
Residual DNase	Ambion DNaseAlert	< LOQ	Pass
Residual RNase	Ambion RNaseAlert	< LOQ	Pass
Appearance	Visual inspection	Clear, colorless	Pass
Residual Host Protein	Cygnus <i>E. coli</i> HCP ELISA Kit	≤ 200 ng/mL	Pass
Residual Host GenomicDNA	Quantitative PCR with Proteinase K digestion	< 1% (w/w)	Pass
Activity Assay, <i>in vitro</i>	Cas9-mediated cleavage of reference DNA	≥80% of uncut reference DNA converted to cleavage products	Pass
Bioburden	Extraction and filtration	< 2 CFU/Sample	Pass
Mycoplasma	Quantitative PCR	Negative for the presence of Mycoplasma	Pass

Quality Control:

Date:

16-AUG-2018

Aldevron Certificate of Analysis Lot MPM020-04

Date of Manufacture: 15-APR-2018



5602 Research Park Blvd
 Madison, Wisconsin 53719, USA
 phone: 1 (701) 297-9256
 toll-free: 1 (877) 787-3362
 fax: 1 (608) 441-3460
 email: protein@aldevron.com

Protein Name:	Item Code:
sNLS-SpCas9-sNLS Nuclease	9212-5MG
Storage Temperature:	Final Buffer:
-10 to -30 °C	25 mM Tris-HCl, 300 mM NaCl, 0.1 mM EDTA, 50% (v/v) Glycerol, pH 7.4
Container Fill:	Usage:
5.0 mg (0.50 mL)	Research use only

Assay	Method	Specifications	Results
Purity, RP-HPLC	RP-HPLC method	≥95%	Pass
Purity, SDS-PAGE	Densitometry SYPRO® Orange	≥95%	Pass
Aggregation	Analytical SEC	≤5% aggregate	Pass
Concentration	UV VIS 280 nm	10.0 mg/mL +/- 10%	Pass
Endotoxin	Kinetic turbidimetric LAL	≤ 10 EU/mg	Pass
Residual DNase	Ambion DNaseAlert	< LOQ	Pass
Residual RNase	Ambion RNaseAlert	< LOQ	Pass
Appearance	Visual inspection	Clear, colorless	Pass
Residual Host Protein	Cygnus <i>E. coli</i> HCP ELISA Kit	≤ 200 ng/mL	Pass
Residual Host GenomicDNA	Quantitative PCR with Proteinase K digestion	< 1% (w/w)	Pass
Activity Assay, <i>in vitro</i>	Cas9-mediated cleavage of reference DNA	≥80% of uncut reference DNA converted to cleavage products	Pass
Bioburden	Extraction and filtration	< 2 CFU/Sample	Pass
Mycoplasma	Quantitative PCR	Negative for the presence of Mycoplasma	Pass

Quality Control:

Date:

20-JUL-2018